

## Department of Bacteriology

### Technical Specifications for PCR Hood

**Quantity – 1 No**

<b>S.No</b>	<b>Technical specification</b>	<b>Complied (yes or No)</b>	<b>If compliance given, mention the page no. of your technical document where the details are provided</b>
1	Microprocessor controller		
2	No installation required – Should be plugged directly into a 220V electrical outlet		
3	ISO 5 vertical laminar flow air		
4	No ductwork required		
5	UV shelf with built-in pipettor holder		
6	Lab event timer		
7	Size: Approx 48 x 24 x 42 Inchs		
8	Clear polycarbonate shell for 360° visibility		
9	Extremely quiet operation		
10	UV and fluorescent lights		
11	Digital UV light timer :01-59 minutes		
12	Access port for electrical cords		